



CERTIFICATE OF TYPE APPROVAL


This is to certify that

Lloyd's Register did undertake the relevant type approval procedures of the equipment detailed below which was found to be in compliance with the essential Pollution Prevention requirements for use on ships and offshore installations classed with Lloyd's Register, and for use on ships and offshore installations when authorised by contracting governments to issue the relevant certificates, licences, permits, etc.

Manufacturer	MMC International Corp.
Address	60 Inip Drive Inwood New York 11096 United States of America (USA)
Type	OIL/WATER INTERFACE DETECTORS (POLLUTION PREVENTION)
Description	Portable Oil/Water Interface Detector Types: "D 2401-2S; D-2401-2UIS and D-2401-2UTS"
Specified Standard	IMO Resolution MEPC.5 (XIII)

The attached Design Appraisal Document forms part of this certificate.

This certificate remains valid unless cancelled or revoked, provided the conditions in the attached Design Appraisal Document are complied with and the equipment remains satisfactory in service.

Date of issue	21 September 2019	Expiry date	20 September 2024
Certificate No.	SAS P190009	Signed	
Sheet No	1 of 3	Name	N. Siddiqui Surveyor to Lloyd's Register EMEA A Member of the Lloyd's Register Group

Note:

This certificate is not valid for equipment, the design or manufacture of which has been varied or modified from the specimen tested. The manufacturer should notify Lloyd's Register of any modification or changes to the equipment in order to obtain a valid Certificate.

Lloyd's Register Group Limited, its affiliates and subsidiaries and their respective officers, employees or agents are, individually and collectively, referred to in this clause as the 'Lloyd's Register'. Lloyd's Register assumes no responsibility and shall not be liable to any person for any loss, damage or expense caused by reliance on the information or advice in this document or howsoever provided, unless that person has signed a contract with the relevant Lloyd's Register entity for the provision of this information or advice and in that case any responsibility or liability is exclusively on the terms and conditions set out in that contract.



Page 2 of 3
Document number SAS P190009
Issue number 1

DESIGN APPRAISAL DOCUMENT

Date 21 September 2019	Quote this reference on all future communications UKITSO/MARPOL/TA/NAS/WP36812157
---------------------------	--

ATTACHMENT TO CERTIFICATE OF TYPE APPROVAL No. SAS P190009

This Design Appraisal Document forms part of the Certificate.

APPROVAL DOCUMENTATION:

<u>Document no.</u>	<u>Revision / issue</u>	<u>Title</u>	<u>Date</u>
71097/916	2	Sensing Probe and PCB Assembly	29 Aug 2002
71097/898	2	Circuit Diagram	03 May 1996
3961	-	KDG Mobrey Test Report	30 Oct 1996

EQUIPMENT TESTED:

Type: Portable oil/water Interface Detector
Model: D-2401-2S
Manufactured by: MMC International Corp.

ACCURACY AND RESPONSE TIMES TESTS:

The maximum error between measured and indicated reading was 3 mm.

In the case of lighter oils response time was less than 1 second.

In case of the heavy oil the response time varied between 3 and 4 seconds.

At Ambient Temperature	Fresh water Density (1.000)			Brackish water density (1.012)			Salt water density (1.025)		
	Measured (mm)	Indicated (mm)	Response Time (Seconds)	Measured (mm)	Indicated (mm)	Response Time (Seconds)	Measured (mm)	Indicated (mm)	Response Time (Seconds)
Unleaded Automotive Gasoline	589.5	586	<1	594.5	591	<1	593	590	<1
Light Diesel Oil	596	593	<1	600.5	598	<1	611	609	<1
Arabian Light Crude Oil	694.5	591	<1	596	593	<1	591.5 (20 °C)	588	<1
							547 (53 °C)	544	<1
Residual Fuel	548 (48 °C)	545 (48 °C)	3 to 4	540 (47 °C)	538 (47 °C)	4	538.5 (49°C)	536 (49 °C)	3 to 4



Lloyd's
Register

Lloyd's Register EMEA (Southampton)

Marine & Offshore UK&I

Lloyd's Register Global Technology Centre, Southampton
Boldrewood Innovation Campus, Burgess Road, Southampton,
SO16 7QF
Telephone +44 330 414 1000 / E-mail tass@lr.org

Page	3 of 3
Document number	SAS P190009
Issue number	1

DESIGN APPRAISAL DOCUMENT

Date 21 September 2019	Quote this reference on all future communications UKITSO/MARPOL/TA/NAS/WP36812157
---------------------------	--

ATTACHMENT TO CERTIFICATE OF TYPE APPROVAL No. SAS P190009

TEMPERATURE TEST:

In the Arabian Light Crude, the difference in level detected at 20°C and 53.1°C was small, less than 1.0 mm.

CONTAMINATION TEST:

The indicated interface level was within 1 mm.

The response time was 3 to 4 seconds.

Salinity of water did not affect the perceived interface level.

CONDITIONS OF CERTIFICATION

1. A copy of this certificate shall be carried on board any ship equipped with the above equipment
2. An operation and maintenance manual should be supplied with each system
3. Production items are to be manufactured in accordance with a quality control system which shall be maintained to ensure that items are of the same standard as the approved prototype

PLACE OF PRODUCTION

MMC International Corp.
60 Inip Drive
Inwood
New York 11056
United States of America (USA)



Nouman Ahmad Siddiqui
Senior Specialist
Statutory Discipline Team - MARPOL
UK&I Marine and Offshore
Lloyd's Register EMEA

Supplementary Type Approval Terms and Conditions

This certificate and Design Appraisal Document relates to type approval, it certifies that the prototype(s) of the product(s) referred to herein has/have been found to meet the applicable design criteria for the use specified herein, it does not mean or imply approval for any other use, nor approval of any products designed or manufactured otherwise than in strict conformity with the said prototype(s).